


Laboratory Accreditation Programmes

Schedule to CERTIFICATE OF ACCREDITATION	
Laboratory	EMC Technologies (NZ) Ltd
Address	47 Mackelvie Street, Grey Lynn, Auckland, 1021
Telephone	09 360-0862
Fax	09 360-0861
URL	www.emctech.com.au
Authorised Representative	Mr Andrew Cutler General Manager
Client No.	2218
Programme	Electrical Testing Laboratory
Accreditation Number	424
Initial Accreditation Date	24 June 1991
Conformance Standard	NZS ISO/IEC 17025:2005 General requirements for the competence of testing and calibration laboratories
Testing Services Summary	3.41 Radiocommunications Equipment 3.42 Electromagnetic Compatibility Testing
Signatories	Mr Andrew Cutler 3.41, 3.42

Authorised: General Manager		Issue 45	Date: 15/08/18	Page 1 of 7
--------------------------------	-------------------------------------------------------------------------------------	----------	----------------	-------------

Schedule to

CERTIFICATE OF ACCREDITATION

EMC Technologies (NZ) Ltd
 Electrical Testing Laboratory
SCOPE OF ACCREDITATION

Accreditation No 424

3.41 Radiocommunications Equipment

- (a) Receiving equipment
- (b) Transmitting equipment


The following tests on analogue AM and FM, VHF and UHF transmitters and receivers in accordance with In-house methods and the methods specified below. Measurements are made over the ranges given.

- i) Transmitter power
- ii) Frequency stability
- iii) Case radiation
- iv) Spurious emissions
- v) Effective radiated power
- vi) Frequency deviation
- vii) Adjacent channel power
- viii) Intermodulation attenuation
- ix) Co-channel rejection
- x) Adjacent channel selectivity
- xi) Intermodulation response
- xii) Blocking
- xiii) Modulation distortion
- xiv) Transmitter transient performance
- xv) Emission masks

Methods

ANSI C63.10-2013	American National Standard of Procedures for Compliance Testing of Unlicensed Wireless Devices (limited to maximum 1 GHz)
ANSI C63.26-2016	Standard for Compliance Testing of Transmitters Used in Licensed Radio Services
ANSI/TIA-603-D	Land Mobile FM or PM Communications Equipment Measurement and Performance Standards
AS/NZS 4268	Radio equipment and systems – Short range devices – Limits and methods of measurement
AS/NZS 4295	VHF/UHF Analogue Speech Equipment
AS/NZS 4415	VHF Maritime Mobile Equipment
AS/NZS 4365	UHF Citizen Band radio and Personal Radio Service
AS/NZS 4768.1	Digital radio equipment operating in land mobile and fixed services bands in the frequency range 29.7 MHz to 1 GHz - Radiofrequency requirements
AS/NZS 4769.1	Radiocommunications equipment used in the paging service - Angle modulated equipment
ETSI EN	Electromagnetic compatibility and Radio spectrum Matters (ERM)
ETSI EN 300 086-1	Land Mobile Service; Radio equipment with an internal or external RF connector intended primarily for analogue speech; Part 1: Technical characteristics and methods of measurement

Authorised:
 General Manager



Issue 45

Date: 15/08/18

Page 2 of 7

Schedule to

CERTIFICATE OF ACCREDITATION

EMC Technologies (NZ) Ltd
 Electrical Testing Laboratory
SCOPE OF ACCREDITATION

Accreditation No 424

ETSI EN 300 113-1	Land mobile service; Radio equipment intended for the transmission of data (and/or speech) using constant or non-constant envelope modulation and having an antenna connector; Part 1: Technical characteristics and methods of measurement
ETSI EN 300 220-1	Short Range Devices (SRD); Radio equipment to be used in the 25 MHz to 1 000 MHz frequency range with power levels ranging up to 500 mW; Part 1: Technical characteristics and test methods
ETSI EN 300 330	Short Range Devices (SRD); Technical characteristics and test methods for radio equipment in the frequency range 9 kHz to 25 MHz and inductive loop systems in the frequency range 9 kHz to 30 MHz
ETSI EN 300 440-1	Short Range Devices (SRD); Radio equipment to be used in the 1 GHz to 40 GHz frequency range; Part 1: Technical characteristics and test methods
FCC 47 CFR FCC MP-5	Parts; 22, 87, 90, 95, 97 & 101 FCC Methods of Measurement of Radio Noise Emissions from Industrial, Scientific and Medical Equipment
RSS 119	Land Mobile and Fixed Ratio Transmitters and Receivers - 27.41 to 960 MHz (i, ii, iii, iv, xiii, xiv and xv above)
RSS 210 RSS 247	Licence-Exempt Radio Apparatus: Category I Equipment Digital Transmission Systems (DTSs), Frequency Hopping Systems (FHSs) and Licence-Exempt Local Area Network (LE-LAN) Devices

Measurement Capabilities

Power (conducted and radiated)
 1 nW to 1000 watts below 1 GHz
 1 nW to 30 watts 1 GHz to 26 GHz

Frequency reference 1 part in 10⁹

Frequency modulation deviation
 10 Hz in 10 kHz
 Temperature – 30 °C to + 60 °C

3.42 Electromagnetic Compatibility Testing


Determination of radio interference and immunity characteristics of industrial, scientific and medical radio frequency equipment, sound and television receivers, household appliances and similar apparatus, fluorescent lamps and luminaires and information technology equipment.

(a) Radiated emissions

Radiated Electric Field Emissions (incl. Low power tx & case radiation)

25 MHz to 1 GHz 3 m and 10 m measurement distances AC mains single and 3 phase	ANSI C63.4-2014 IEC, AS/NZS CISPR 11 IEC, AS/NZS CISPR 12 IEC, AS/NZS CISPR 13
-----------------------------------------------------------------------------------------	-----------------------------------------------------------------------------------------

Authorised:
 General Manager



Issue 45

Date: 15/08/18

Page 3 of 7

Schedule to

CERTIFICATE OF ACCREDITATION

EMC Technologies (NZ) Ltd
 Electrical Testing Laboratory
SCOPE OF ACCREDITATION

Accreditation No 424

supply up to 16 amps

- IEC, AS/NZS CISPR 14
- IEC, AS/NZS CISPR 16-1
- IEC, AS/NZS CISPR 16-2
- IEC, AS/NZS CISPR 32
- AS/NZS 3200.1.2
- AS/NZS 4268
- EN 55011
- EN 55032
- EN 60945, clauses 4.5.1, 9, 10 only
- EN / IEC 60601-1-2
- EN / IEC 61000-6-3
- EN / IEC 61000-6-4
- EN 300 086
- EN 300 220-1
- EN 301 489-1 / EN 301 489-3
- EN 301 489-1 / EN 301 489-4
- EN 301 489-1 / EN 301 489-5

Radiated Magnetic Field Emissions (including Van Veen Loop measurements)

9 kHz to 30 MHz

- IEC, AS/NZS CISPR 11
- IEC, AS/NZS CISPR 15
- AS/NZS 3200.1.2
- EN 55011
- EN 55015
- EN / IEC 60601-1-2
- EN 60945, clauses 4.5.1, 9, 10 only
- EN 300 330

Note: All radiated emission measurements are carried out at the laboratory's open area test site (OATS).


(b) Radiated immunity

Radiated Electromagnetic Immunity Testing

27 MHz to 2800 MHz
 1 V/m, 3 V/m, 10 V/m
 Radiated field area 0.5 m x 0.5 m
 maximum

- IEC, AS/NZS CISPR 14-2
- IEC, AS/NZS CISPR 24
- AS/NZS 3200.1.2
- EN 55024
- IEC AS/NZS 60335.1 clause 19.11.4.2
- EN 60601-1-2/ IEC 60601-1-2
- EN 60945, clauses 4.5.1, 9, 10 only
- IEC / EN 61000-4-3
- EN 61547
- EN 301 489-1/ EN 301 489-3
- EN 301 489-1/ EN 301 489-4
- EN 301 489-1/ EN 301 489-5

Authorised:
 General Manager



Issue 45

Date: 15/08/18

Page 4 of 7

Schedule to

CERTIFICATE OF ACCREDITATION

EMC Technologies (NZ) Ltd
 Electrical Testing Laboratory
SCOPE OF ACCREDITATION

Accreditation No 424

	EN 55014 IEC 61000-6-3 IEC 61000-6-4
Harmonic current emissions DC to 40 th harmonic 240 Vac, up to 16 A per phase	EN IEC 61000-3-2
Voltage changes, fluctuations and flicker 240 Vac, up to 16 A per phase	EN IEC 61000-3-3
(d) Conducted immunity	
Electrostatic Discharge Up to 15 kV air discharge	IEC, AS/NZS CISPR 14-2 IEC, AS/NZS CISPR 24 AS/NZS 3200.1.2 EN 55014-2 EN 55024 IEC AS/NZS 60335.1 clause 19.11.4.1 EN / IEC 60601-1-2 EN 60945, clauses 4.5.1, 9, 10 only IEC 61000-4-2 EN 301 489-1 / EN 301 489-3 EN 301 489 1 / EN 304 489-4 EN 301 489 1 / EN 304 489-5
Up to 8kV contact discharge	As above plus EN 61547
Electrical Fast Transients AC mains up to 240 V single phase, 16 amps Power port and IO, single and control port available Voltage 200 V to 4000 V +ve & -ve tBURST 0.1 mS to 20 mS fBURST (0.1 to 2.5, 5, 100) kHz > 2kV	IEC, AS/NZS CISPR 14-2 IEC, AS/NZS CISPR 24 AS/ NZS 3200.1.2 EN 55014-2 EN 55024 IEC AS/NZS 60335.1 clause 19.11.4.3 EN 60601-1-2/ IEC 60601-1-2 EN 60945, clauses 4.5.1, 9, 10 only IEC 61000-4-4 EN 61547 EN 301 489-1 / EN 301 489-3 EN 301 489-1 / EN 301 489-4 EN 301 489-1 / EN 301 489-5
Surges 0.5 kV to 2.0 kV (line to line) 0.5 kV to 4.0 kV (line to ground)	IEC, AS/NZS CISPR 14-2 IEC, AS/NZS CISPR 24 AS/NZS 3200.1.2 EN 55014-2

Authorised: General Manager <i>P. Bone</i>	Issue 45	Date: 15/08/18	Page 6 of 7
-----------------------------------------------	----------	----------------	-------------


Schedule to

CERTIFICATE OF ACCREDITATION

EMC Technologies (NZ) Ltd
 Electrical Testing Laboratory
SCOPE OF ACCREDITATION

Accreditation No 424

<p>Voltage dips and short interruptions DC Current Max 24 A AC Voltage Max 280 Vrms</p>	<p>EN 55024 IEC AS/NZS 60335.1 clause 19.11.4.4 EN / IEC 60601-1-2 EN 60945, clauses 4.5.1, 9, 10 only IEC 61000-4-5 EN 61547 EN 301 489-1/ EN 301 489-3 EN 301 489-1/ EN 301 489-4 EN 301 489-1/ EN 301 489-5</p>
<p>Radio frequency injected currents 1 V, 3 V, 10 V</p>	<p>IEC, AS/NZS CISPR 14-2 IEC, AS/NZS CISPR 24 EN 55014-2 EN 55024 IEC AS/NZS 60335.1 clause 19.11.4.6 EN 60601-1-2/IEC 60601-1-2 EN 60945, clauses 4.5.1, 9, 10 only IEC 61000-4-11 EN 61547 EN 301 489-1/ EN 301 489-3 EN 301 489-1/ EN 301 489-4 EN 301 489-1/ EN 301 489-5</p>
<p>Harmonics and interharmonics</p>	<p>IEC, AS/NZS CISPR 14-2 IEC, AS/NZS CISPR 24 EN 55014-2 EN 55024 IEC AS/NZS 60335.1 clause 19.11.4.5 EN / IEC 60601-1-2 EN 60945, clauses 4.5.1, 9, 10 only IEC 61000-4-6 EN 301 489-1/ EN 301 489-3 EN 301 489-1/ EN 301 489-4 EN 301 489-1/ EN 301 489-5</p>
<p>Harmonics and interharmonics</p>	<p>IEC AS/NZS 60335.1 clause 19.11.4.7 IEC/EN 61000-4-13</p>

Authorised: General Manager 	Issue 45	Date: 15/08/18	Page 7 of 7
--------------------------------------------------------------------------------------------------------------------	----------	----------------	-------------